



\*<sup>1</sup> Slight vignetting may occur in the periphery of the field of view in combination with an additional intermediate attachment. \*<sup>2</sup> Slight vignetting may occur in the periphery of the field of view in combination with fluorescence illuminator.  
\*<sup>3</sup> Field of view is limited when using more than two intermediate tubes. \*<sup>4</sup> Sub port can accept U-TV1x and 2/3" or less CCDs. \*<sup>5</sup> F.N. is 22 when fluorescence observation.